

# NABL Assessment Readiness Audit

Diagnostics, Labs & Imaging

Document type: Plan / SOP / strategy / guide · Generated 14 Oct 2026

This guide explains what the tool produces, how to use the output, and what to watch for. The actual output (draft/report) is generated on the tool page.

## Background

Runs a readiness check for a NABL assessment — quality system, competence, QC, records and common non-conformities — with a gap list and fixes. Use it before the assessor visits. For the quality manual use the NABL Quality Manual Builder.

## Purpose of this tool & output

Check if you are ready for the NABL assessor

## How the output is used

Use it to check if you are ready for the NABL assessor — a self-audit against the standard that finds gaps before the assessment.

## Important cautions

Assessors check implementation and records over time (IQC, EQAS, audits, reviews), not just documents — a quality system needs to have been running for months. This is a self-check; fix gaps genuinely rather than papering over them. Align to current NABL 112 and ISO 15189.

## Companion documents

- NABL Quality Manual Builder
- Internal Quality Control (IQC) Plan
- External Quality Assurance (EQAS/PT) Plan
- Standard SOP Library Builder
- Non-conformity & CAPA Log
- Calibration & Maintenance Schedule

This is AI-assisted guidance. Verify with a qualified professional before any binding decision. Fees, authorities and processes vary by state and change over time — confirm with your state authority.